
Midmark M11 - v1.0.0
Total Cycles: 4003

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2023 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.5 F 0.0
2:00 148.6 F 2.8
4:00 168.9 F 6.3
6:00 212.2 F 11.0
8:00 220.2 F 4.5
10:00 226.6 F 9.5
12:00 236.6 F 9.1
14:00 241.9 F 11.7
16:00 239.6 F 10.2
18:00 253.2 F 16.9
20:00 264.8 F 23.7
22:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.1 F 28.7
2:00 272.4 F 29.0
3:00 272.0 F 29.1
4:00 271.9 F 28.7

Min 271.4 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 231.7 F 0.0
10:00 207.0 F 0.0
15:00 189.6 F 0.0
20:00 186.4 F 0.0
25:00 182.9 F 0.0
30:00 185.4 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

11/09/2023 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____