
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005913
8:02 06/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:30
STERILIZING 3:39
136.5 C 223kPa 3:39
136.3 C 223kPa 4:09
136.1 C 221kPa 4:39
136.4 C 223kPa 5:09
136.4 C 223kPa 5:39
136.4 C 221kPa 6:09
136.4 C 223kPa 6:39
136.4 C 222kPa 7:09
136.4 C 222kPa 7:09
VENTING 7:10
AIR DRYING 8:03
CYCLE COMPLETE 10:43

Digital Signature #
0F6161830ECA5851
