

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 003537  
16:06 26/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:03  
STERILIZING 4:05  
136.4 C 223kPa 4:05  
136.2 C 224kPa 4:35  
136.1 C 221kPa 5:05  
136.3 C 223kPa 5:35  
136.5 C 224kPa 6:05  
136.3 C 224kPa 6:35  
136.2 C 223kPa 7:05  
136.3 C 223kPa 7:35  
136.3 C 223kPa 7:35  
VENTING 7:35  
AIR DRYING 8:31  
CYCLE COMPLETE 8:41  
-----

Digital Signature #  
F229461C22006ECB  
-----