

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2749

Name: RITTER2  
Sterilizer ID: V2313228  
08/10/2023 09:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 187.7 F 1.8  
4:00 192.9 F 6.8  
6:00 215.6 F 12.5  
8:00 225.5 F 8.1  
10:00 240.1 F 12.6  
12:00 239.6 F 11.6  
14:00 239.3 F 10.6  
16:00 252.9 F 17.5  
18:00 264.6 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.5 F 29.6  
2:00 271.7 F 29.2  
3:00 272.5 F 29.6  
4:00 271.5 F 29.1  
  
Min 271.5 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 223.6 F 0.0  
10:00 229.1 F 0.0  
15:00 229.4 F 0.0  
20:00 227.3 F 0.0  
25:00 225.7 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 19:52  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

08/10/2023 10:08 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_