
Midmark M11 - v1.0.5
Total Cycles: 1260

Name: RITTER2
Sterilizer ID: V2313228
07/06/2022 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 157.6 F 2.2
4:00 166.2 F 5.3
6:00 183.8 F 9.7
8:00 208.5 F 7.4
10:00 220.0 F 8.5
12:00 233.4 F 13.5
14:00 236.7 F 10.4
16:00 237.2 F 8.4
18:00 248.0 F 14.9
20:00 259.5 F 21.4
22:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 30.0
2:00 271.8 F 29.4
3:00 271.5 F 29.7
4:00 272.7 F 30.0

Min 271.2 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 207.0 F 0.0
10:00 224.0 F 0.0
15:00 223.5 F 0.0
20:00 222.8 F 0.0
25:00 225.2 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

07/06/2022 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____