

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4083

Name: RITTER 1  
Sterilizer ID: V2178198  
12/05/2023 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	162.9 F	0.0
2:00	160.7 F	1.5
4:00	160.3 F	4.5
6:00	184.0 F	8.2
8:00	202.1 F	12.4
10:00	214.5 F	4.8
12:00	227.0 F	8.9
14:00	237.0 F	13.4
16:00	236.5 F	10.0
18:00	237.5 F	9.0
20:00	246.9 F	13.6
22:00	257.9 F	19.4
24:00	267.9 F	25.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.1 F	28.3
1:00	272.4 F	28.9
2:00	271.2 F	28.2
3:00	271.9 F	29.0
4:00	273.1 F	29.3

Min 270.9 F 28.1  
Max 273.2 F 29.5

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	191.7 F	0.0
5:00	232.9 F	0.0
10:00	214.0 F	0.0
15:00	208.9 F	0.0
20:00	206.6 F	0.0
25:00	206.8 F	0.0
30:00	203.2 F	0.0

Drying Complete

FILLING: 1:02  
HEATING: 25:24  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:04:08  
CYCLE COMPLETE

12/05/2023 11:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_