
Midmark M11 - v1.0.0
Total Cycles: 3457

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 195.7 F 2.9
4:00 204.3 F 7.7
6:00 223.0 F 12.4
8:00 230.8 F 9.2
10:00 238.8 F 9.4
12:00 244.1 F 13.0
14:00 243.2 F 12.1
16:00 256.8 F 19.1
18:00 268.6 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.8 F 28.8
3:00 272.3 F 28.9
4:00 271.5 F 28.6

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 226.6 F 0.0
10:00 235.0 F 0.0
15:00 235.2 F 0.0
20:00 226.3 F 0.0
25:00 223.5 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

06/28/2023 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____