

-----  
Midmark M11 - v1.0.5  
Total Cycles: 546

Name: RITTER2  
Sterilizer ID: V2313228  
11/11/2021 03:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
2:00 182.0 F 4.2  
4:00 204.0 F 10.5  
6:00 221.6 F 6.9  
8:00 237.1 F 13.5  
10:00 241.7 F 12.3  
12:00 242.4 F 11.9  
14:00 256.8 F 19.3  
16:00 269.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.4 F 29.3  
2:00 272.3 F 29.7  
3:00 271.6 F 29.2  
4:00 272.2 F 29.2  
  
Min 271.4 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 247.7 F 0.0  
10:00 221.2 F 0.0  
15:00 196.9 F 0.0  
20:00 192.7 F 0.0  
25:00 191.0 F 0.0  
30:00 196.8 F 0.0  
Drying Complete

FILLING: 2:40  
HEATING: 16:56  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 00:56:34  
CYCLE COMPLETE

11/11/2021 04:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_