
Midmark M11 - v1.0.0
Total Cycles: 3619

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2023 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.2 F 0.0
2:00 119.2 F 0.0
4:00 134.6 F 0.0
6:00 174.9 F 2.5
8:00 193.1 F 6.1
10:00 204.6 F 9.9
12:00 217.3 F 12.1
14:00 223.5 F 6.9
16:00 234.5 F 11.2
18:00 233.0 F 7.6
20:00 244.2 F 13.0
22:00 242.3 F 11.4
24:00 254.0 F 17.3
26:00 264.4 F 23.4
28:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.3 F 28.0
3:00 272.5 F 28.9
4:00 271.9 F 28.4

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 210.6 F 0.0
10:00 205.5 F 0.0
15:00 214.0 F 0.0
20:00 211.9 F 0.0
25:00 213.0 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:32
CYCLE COMPLETE

08/04/2023 04:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____