
Midmark M11 - v1.0.5
Total Cycles: 1811

Name: RITTER2
Sterilizer ID: V2313228
12/06/2022 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 189.0 F 3.7
4:00 194.9 F 9.1
6:00 215.1 F 6.8
8:00 231.7 F 11.1
10:00 234.1 F 9.0
12:00 238.9 F 9.6
14:00 249.6 F 15.5
16:00 262.5 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.5 F 29.7
2:00 271.6 F 29.1
3:00 272.6 F 29.6
4:00 271.6 F 29.0

Min 271.6 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 220.1 F 0.0
10:00 230.0 F 0.0
15:00 232.1 F 0.0
20:00 229.6 F 0.0
25:00 227.0 F 0.0
30:00 226.2 F 0.0

Drying Complete

FILLING: 2:15
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

12/06/2022 12:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____