
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000760
14:53 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:54
STERILIZING 3:58
136.0 C 228kPa 3:58
136.4 C 233kPa 4:28
136.3 C 232kPa 4:58
136.4 C 233kPa 5:28
136.0 C 230kPa 5:58
136.4 C 233kPa 6:28
136.2 C 232kPa 6:58
136.4 C 232kPa 7:28
136.4 C 232kPa 7:28
VENTING 7:28
AIR DRYING 8:22
CYCLE COMPLETE 8:23

Digital Signature #
25AEB3A77D80D016
