

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005005  
16:11 24/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:57  
STERILIZING 3:06  
136.4 C 220kPa 3:06  
136.4 C 224kPa 3:36  
136.3 C 221kPa 4:06  
136.4 C 224kPa 4:36  
136.4 C 224kPa 5:06  
136.0 C 221kPa 5:36  
136.4 C 225kPa 6:06  
136.4 C 224kPa 6:36  
136.4 C 224kPa 6:36  
VENTING 6:36  
AIR DRYING 7:24  
CYCLE COMPLETE 11:31

-----

Digital Signature #  
1CE3774961CB79C7  
-----