
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003043
14:34 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:45
STERILIZING 4:50
136.5 C 220kPa 4:50
136.3 C 224kPa 5:20
136.2 C 223kPa 5:50
136.4 C 223kPa 6:20
136.2 C 220kPa 6:50
136.1 C 222kPa 7:20
136.4 C 222kPa 7:50
136.1 C 221kPa 8:20
136.1 C 221kPa 8:20
VENTING 8:20
AIR DRYING 9:17
CYCLE COMPLETE 9:30

Digital Signature #
216BE39EA37E33A7
