
Midmark M11 - v1.0.0
Total Cycles: 1495

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2022 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 78.9 F 0.0
4:00 106.3 F 0.1
6:00 124.0 F 0.1
8:00 150.1 F 0.7
10:00 172.3 F 2.3
12:00 193.2 F 3.8
14:00 208.0 F 6.3
16:00 215.0 F 9.4
18:00 225.5 F 12.7
20:00 228.1 F 7.2
22:00 239.2 F 11.5
24:00 234.4 F 8.2
26:00 245.1 F 12.5
28:00 244.5 F 12.3
30:00 256.0 F 18.3
32:00 266.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.8
2:00 271.6 F 28.0
3:00 272.1 F 28.7
4:00 272.1 F 28.5

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 195.5 F 0.0
10:00 193.9 F 0.0
15:00 218.8 F 0.0
20:00 224.3 F 0.0
25:00 229.5 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 0:36
HEATING: 33:47
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:11:43
CYCLE COMPLETE

01/05/2022 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____