
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002505
14:42 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:26
STERILIZING 3:34
136.0 C 222kPa 3:34
136.4 C 221kPa 4:05
136.4 C 226kPa 4:35
136.2 C 224kPa 5:05
136.3 C 225kPa 5:35
136.4 C 225kPa 6:05
136.4 C 226kPa 6:35
136.4 C 227kPa 7:05
136.4 C 227kPa 7:05
VENTING 7:05
AIR DRYING 8:01
CYCLE COMPLETE 25:31

Digital Signature #
065E6434D963B46C
