
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002573
7:44 14/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:37
STERILIZING 3:38
136.3 C 222kPa 3:38
135.6 C 220kPa 4:08
136.4 C 225kPa 4:38
136.4 C 225kPa 5:08
136.4 C 224kPa 5:38
136.3 C 223kPa 6:08
136.4 C 223kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:08
VENTING 7:08
AIR DRYING 7:58
CYCLE COMPLETE 25:42

Digital Signature #
D007060D808BDBB1
