
Midmark M11 - v1.0.5
Total Cycles: 2940

Name: RITTER2
Sterilizer ID: V2313228
09/27/2023 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.3 F 0.0
2:00 147.6 F 0.0
4:00 178.3 F 2.9
6:00 200.9 F 7.8
8:00 218.1 F 13.1
10:00 232.0 F 8.1
12:00 243.1 F 13.4
14:00 240.8 F 11.2
16:00 239.1 F 9.9
18:00 253.0 F 16.8
20:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.4 F 28.8
2:00 272.4 F 29.2
3:00 271.8 F 28.3
4:00 272.7 F 29.1

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 205.6 F 0.0
10:00 188.7 F 0.0
15:00 186.0 F 0.0
20:00 183.8 F 0.0
25:00 181.5 F 0.0
30:00 180.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

09/27/2023 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____