

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3459

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2023 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 186.0 F 3.4  
4:00 193.3 F 7.9  
6:00 218.2 F 13.2  
8:00 228.8 F 8.0  
10:00 241.1 F 13.3  
12:00 239.7 F 10.9  
14:00 237.6 F 9.4  
16:00 252.0 F 16.3  
18:00 264.0 F 23.3  
20:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.5 F 29.1  
2:00 271.7 F 28.4  
3:00 271.6 F 28.7  
4:00 272.5 F 29.1

Min 271.2 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 226.1 F 0.0  
10:00 233.6 F 0.0  
15:00 229.7 F 0.0  
20:00 230.4 F 0.0  
25:00 227.9 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:02  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

06/28/2023 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_