
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000675
13:46 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:41
STERILIZING 3:46
136.0 C 228kPa 3:46
136.4 C 233kPa 4:16
136.2 C 232kPa 4:46
136.4 C 233kPa 5:16
136.4 C 233kPa 5:46
136.5 C 233kPa 6:16
136.2 C 231kPa 6:46
136.4 C 233kPa 7:16
136.4 C 233kPa 7:16
VENTING 7:16
AIR DRYING 8:12
CYCLE COMPLETE 8:13

Digital Signature #
0663BFDC0AECAD24
