
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006532
8:29 30/10/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 2:00
STERILIZING 2:57
136.4 C 222kPa 2:57
136.5 C 224kPa 3:27
136.4 C 223kPa 3:57
136.4 C 224kPa 4:27
136.4 C 223kPa 4:57
136.4 C 223kPa 5:27
136.4 C 221kPa 5:57
136.4 C 220kPa 6:27
136.4 C 220kPa 6:27
VENTING 6:27
AIR DRYING 7:17
CYCLE COMPLETE 7:49

Digital Signature #
42298F3D4FA9653C
