
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001709
10:41 07/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:03
STERILIZING 4:04
136.5 C 223kPa 4:04
136.2 C 223kPa 4:34
136.5 C 224kPa 5:04
136.4 C 222kPa 5:34
135.9 C 221kPa 6:04
136.3 C 223kPa 6:34
136.4 C 225kPa 7:04
136.4 C 224kPa 7:34
136.4 C 224kPa 7:34
VENTING 7:34
AIR DRYING 8:21
CYCLE COMPLETE 8:27

Digital Signature #
7EBC1654A8277359
