

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1250

Name: RITTER 1  
Sterilizer ID: V2178198  
10/08/2021 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.9 F 0.0  
2:00 156.3 F 0.0  
4:00 177.3 F 2.7  
6:00 192.9 F 6.4  
8:00 209.8 F 10.9  
10:00 225.0 F 5.1  
12:00 233.0 F 9.9  
14:00 236.6 F 8.8  
16:00 242.4 F 12.2  
18:00 240.7 F 10.4  
20:00 253.3 F 16.8  
22:00 264.2 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.1 F 29.2  
2:00 272.0 F 28.8  
3:00 272.7 F 29.4  
4:00 271.6 F 28.6

Min 271.0 F 28.3  
Max 273.3 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 221.3 F 0.0  
10:00 199.0 F 0.0  
15:00 197.5 F 0.0  
20:00 195.6 F 0.0  
25:00 230.5 F 0.0  
30:00 237.9 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

10/08/2021 02:15 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02