
Midmark M11 - v1.0.5
Total Cycles: 1210

Name: RITTER2
Sterilizer ID: V2313228
06/22/2022 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 139.2 F 1.6
4:00 143.6 F 4.0
6:00 167.5 F 7.6
8:00 188.0 F 11.9
10:00 209.9 F 5.0
12:00 223.1 F 9.7
14:00 231.5 F 9.4
16:00 239.0 F 11.6
18:00 237.7 F 9.9
20:00 250.7 F 16.2
22:00 262.2 F 22.7
24:00 272.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.7 F 29.9

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 202.1 F 0.0
10:00 208.1 F 0.0
15:00 207.9 F 0.0
20:00 206.2 F 0.0
25:00 205.7 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

06/22/2022 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____