
Midmark M11 - v1.0.5
Total Cycles: 1317

Name: RITTER2
Sterilizer ID: V2313228
07/19/2022 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 148.2 F 0.0
4:00 168.0 F 0.5
6:00 200.3 F 5.1
8:00 226.1 F 10.1
10:00 227.8 F 7.3
12:00 237.5 F 11.4
14:00 235.6 F 9.2
16:00 238.9 F 9.7
18:00 248.2 F 14.8
20:00 260.1 F 21.5
22:00 270.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.4
2:00 271.8 F 29.4
3:00 272.4 F 29.3
4:00 271.8 F 29.4

Min 271.4 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.5 F 0.0
5:00 221.8 F 0.0
10:00 224.7 F 0.0
15:00 204.2 F 0.0
20:00 206.1 F 0.0
25:00 208.9 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

07/19/2022 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____