
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007732
8:12 13/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:15
136.5 C 221kPa 4:15
136.2 C 219kPa 4:45
136.3 C 218kPa 5:15
136.4 C 221kPa 5:45
136.2 C 222kPa 6:15
136.4 C 222kPa 6:45
136.3 C 220kPa 7:15
136.4 C 223kPa 7:45
136.4 C 223kPa 7:45
VENTING 7:45
AIR DRYING 8:39
CYCLE COMPLETE 11:56

Digital Signature #
9296C832E1D22410
