

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001788  
12:43 12/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:30  
STERILIZING 3:28  
136.0 C 220kPa 3:28  
136.2 C 222kPa 3:58  
136.4 C 225kPa 4:28  
136.0 C 222kPa 4:58  
136.4 C 225kPa 5:28  
136.4 C 225kPa 5:58  
136.4 C 225kPa 6:28  
136.4 C 224kPa 6:58  
136.4 C 224kPa 6:58  
VENTING 6:58  
AIR DRYING 7:47  
CYCLE COMPLETE 37:47  
-----

Digital Signature #  
42CCD6AEC3C79F84  
-----