

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1830

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2022 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
2:00 198.8 F 3.4  
4:00 208.4 F 9.2  
6:00 223.9 F 7.0  
8:00 234.6 F 11.2  
10:00 237.0 F 9.3  
12:00 240.1 F 9.3  
14:00 250.2 F 16.0  
16:00 262.9 F 23.3  
18:00 271.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.7 F 29.6  
2:00 271.8 F 29.2  
3:00 272.8 F 29.7  
4:00 272.1 F 29.1  
  
Min 271.8 F 29.0  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 205.7 F 0.0  
10:00 218.8 F 0.0  
15:00 226.6 F 0.0  
20:00 229.5 F 0.0  
25:00 228.8 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 2:41  
HEATING: 18:03  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

12/13/2022 11:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_