
Midmark M11 - v1.0.0
Total Cycles: 3015

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2023 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.8 F 0.0
2:00 166.9 F 0.6
4:00 188.5 F 2.6
6:00 203.8 F 5.7
8:00 224.8 F 10.4
10:00 221.5 F 5.4
12:00 232.6 F 10.0
14:00 233.5 F 8.4
16:00 243.2 F 12.7
18:00 241.7 F 11.2
20:00 254.6 F 17.9
22:00 266.1 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.6 F 28.5
3:00 272.3 F 29.4
4:00 272.3 F 29.0

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 214.4 F 0.0
10:00 201.2 F 0.0
15:00 200.2 F 0.0
20:00 229.7 F 0.0
25:00 228.7 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

03/16/2023 03:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____