
Midmark M11 - v1.0.0
Total Cycles: 4576

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2024 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C102: FILL MODE

** ITEMS NOT STERILE **

05/23/2024 12:27 PM