

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1082

Name: RITTER2  
Sterilizer ID: V2313228  
05/09/2022 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.3 F 0.0  
2:00 146.4 F 1.8  
4:00 164.9 F 5.3  
6:00 189.8 F 9.7  
8:00 207.8 F 6.4  
10:00 223.6 F 8.5  
12:00 235.8 F 12.5  
14:00 238.0 F 10.7  
16:00 235.4 F 8.7  
18:00 249.1 F 15.2  
20:00 260.5 F 21.6  
22:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 273.0 F 30.0  
2:00 272.1 F 29.2  
3:00 272.3 F 29.9  
4:00 273.2 F 29.8

Min 271.5 F 29.1  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
5:00 233.9 F 0.0  
10:00 210.2 F 0.0  
15:00 186.7 F 0.0  
20:00 182.8 F 0.0  
25:00 211.5 F 0.0  
30:00 210.5 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:35  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:44  
CYCLE COMPLETE

05/09/2022 10:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_