
Midmark M11 - v1.0.5
Total Cycles: 2270

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 183.0 F 4.4
4:00 204.2 F 9.3
6:00 219.2 F 7.3
8:00 232.2 F 10.9
10:00 233.7 F 8.9
12:00 238.5 F 9.9
14:00 248.9 F 15.2
16:00 261.7 F 22.4
18:00 273.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.8
2:00 271.8 F 29.0
3:00 272.2 F 29.8
4:00 272.5 F 29.6

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 234.8 F 0.0
10:00 221.6 F 0.0
15:00 220.6 F 0.0
20:00 213.6 F 0.0
25:00 214.6 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

04/19/2023 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____