
Midmark M11 - v1.0.0
Total Cycles: 870

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2021 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 168.5 F 1.8
4:00 173.1 F 5.1
6:00 193.5 F 9.0
8:00 210.9 F 13.5
10:00 223.4 F 6.5
12:00 236.3 F 11.1
14:00 232.1 F 7.5
16:00 244.7 F 13.2
18:00 243.0 F 11.7
20:00 255.5 F 18.1
22:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.8 F 28.9
2:00 271.9 F 28.2
3:00 272.0 F 28.6
4:00 272.7 F 28.8

Min 271.5 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 216.3 F 0.0
10:00 217.9 F 0.0
15:00 224.0 F 0.0
20:00 206.7 F 0.0
25:00 194.3 F 0.0
30:00 184.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:41
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

06/08/2021 09:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____