
Midmark M11 - v1.0.0
Total Cycles: 2256

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 193.8 F 0.2
4:00 201.8 F 1.5
6:00 220.2 F 5.8
8:00 233.2 F 10.5
10:00 235.1 F 7.3
12:00 243.0 F 12.9
14:00 241.6 F 11.3
16:00 239.7 F 10.2
18:00 253.7 F 17.1
20:00 265.5 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.7
2:00 271.5 F 28.0
3:00 272.8 F 28.8
4:00 271.9 F 28.1

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 220.8 F 0.0
10:00 235.3 F 0.0
15:00 232.6 F 0.0
20:00 226.8 F 0.0
25:00 225.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

08/29/2022 10:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____