

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3950

Name: RITTER2  
Sterilizer ID: V2313228  
09/18/2024 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.9 F 0.0  
2:00 81.6 F 0.0  
4:00 82.2 F 0.0  
6:00 108.5 F 0.1  
8:00 152.5 F 0.3  
10:00 179.1 F 1.9  
12:00 191.5 F 5.2  
14:00 211.9 F 8.5  
16:00 220.9 F 12.1  
18:00 226.2 F 6.1  
20:00 235.1 F 10.2  
22:00 237.0 F 9.4  
24:00 242.6 F 12.0  
26:00 239.5 F 10.3  
28:00 251.7 F 16.3  
30:00 262.5 F 22.6  
32:00 272.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.6  
1:00 272.2 F 28.9  
2:00 272.5 F 29.1  
3:00 272.5 F 29.4  
4:00 271.8 F 29.2  
  
Min 271.4 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 203.0 F 0.0  
10:00 185.4 F 0.0  
15:00 183.8 F 0.0  
20:00 181.1 F 0.0  
25:00 178.4 F 0.0  
30:00 179.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 32:13  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:10:34  
CYCLE COMPLETE

09/18/2024 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_