

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4052

Name: RITTER 1  
Sterilizer ID: V2178198  
11/22/2023 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.1 F 0.0  
2:00 129.8 F 1.0  
4:00 169.0 F 3.5  
6:00 184.7 F 6.8  
8:00 202.8 F 10.6  
10:00 212.9 F 3.6  
12:00 224.3 F 6.9  
14:00 231.9 F 10.9  
16:00 233.0 F 6.9  
18:00 241.7 F 12.3  
20:00 240.0 F 10.6  
22:00 252.7 F 16.9  
24:00 264.1 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.7 F 29.2  
2:00 271.6 F 28.4  
3:00 272.6 F 29.1  
4:00 271.7 F 28.4

Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 221.0 F 0.0  
10:00 228.3 F 0.0  
15:00 231.7 F 0.0  
20:00 226.7 F 0.0  
25:00 226.6 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 25:52  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:04:24  
CYCLE COMPLETE

11/22/2023 05:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_