
Midmark M11 - v1.0.0
Total Cycles: 860

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2021 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.2 F 0.0
2:00 121.5 F 0.7
4:00 144.2 F 2.2
6:00 171.0 F 4.9
8:00 189.6 F 8.2
10:00 211.2 F 12.0
12:00 217.2 F 5.4
14:00 229.4 F 9.6
16:00 235.4 F 10.3
18:00 240.4 F 11.3
20:00 238.2 F 9.5
22:00 251.0 F 15.7
24:00 262.1 F 22.0
26:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.1
3:00 272.5 F 29.0
4:00 272.2 F 28.5

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 219.8 F 0.0
10:00 214.8 F 0.0
15:00 208.6 F 0.0
20:00 205.4 F 0.0
25:00 203.8 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

06/01/2021 02:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____