
Midmark M11 - v1.0.0
Total Cycles: 1394

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2021 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.7 F 0.0
2:00 154.8 F 1.8
4:00 169.1 F 4.6
6:00 200.9 F 9.1
8:00 219.1 F 7.1
10:00 225.1 F 7.9
12:00 237.4 F 12.7
14:00 236.4 F 9.6
16:00 238.7 F 9.2
18:00 247.9 F 14.2
20:00 259.8 F 20.8
22:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.1 F 28.4
3:00 272.2 F 29.1
4:00 271.4 F 28.5

Min 271.0 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 216.2 F 0.0
10:00 214.1 F 0.0
15:00 209.0 F 0.0
20:00 209.0 F 0.0
25:00 212.6 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

11/24/2021 02:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____