
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004221
15:42 23/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:44
PRESSURIZING 0:51
STERILIZING 1:34
136.0 C 220kPa 1:34
136.4 C 223kPa 2:04
136.0 C 223kPa 2:34
135.9 C 222kPa 3:04
136.4 C 223kPa 3:34
136.4 C 222kPa 4:04
136.4 C 221kPa 4:34
136.1 C 220kPa 5:04
136.1 C 220kPa 5:04
VENTING 5:04
AIR DRYING 5:47
CYCLE COMPLETE 12:01

Digital Signature #
DE62AE5BCEBEE7AE
