
Midmark M11 - v1.0.0
Total Cycles: 882

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.9 F 0.0
2:00 155.6 F 2.1
4:00 172.5 F 5.9
6:00 197.1 F 11.2
8:00 222.5 F 5.8
10:00 227.5 F 11.7
12:00 235.8 F 8.6
14:00 238.2 F 9.7
16:00 246.6 F 13.5
18:00 258.8 F 20.5
20:00 270.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.3
1:00 272.2 F 29.3
2:00 271.9 F 29.3
3:00 272.3 F 29.2
4:00 272.4 F 29.3

Min 271.9 F 29.2
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 227.5 F 0.0
10:00 232.3 F 0.0
15:00 229.8 F 0.0
20:00 236.7 F 0.0
25:00 224.6 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

06/10/2021 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____