
Midmark M11 - v1.0.5
Total Cycles: 1524

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.1 F 0.0
2:00 190.1 F 3.3
4:00 195.7 F 8.2
6:00 213.5 F 13.3
8:00 230.5 F 8.6
10:00 239.8 F 13.3
12:00 239.4 F 11.0
14:00 236.7 F 9.5
16:00 249.9 F 15.8
18:00 260.7 F 22.0
20:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.6
1:00 271.5 F 29.5
2:00 272.2 F 29.6
3:00 272.9 F 29.9
4:00 271.7 F 30.1

Min 271.1 F 29.5
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 229.2 F 0.0
10:00 226.2 F 0.0
15:00 223.6 F 0.0
20:00 218.2 F 0.0
25:00 215.3 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:60
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

09/22/2022 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____