

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2180

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2023 02:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 181.8 F 4.0  
4:00 194.1 F 8.9  
6:00 216.1 F 7.9  
8:00 227.4 F 9.7  
10:00 233.4 F 8.7  
12:00 243.8 F 12.7  
14:00 244.8 F 13.1  
16:00 257.9 F 20.2  
18:00 270.0 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.5  
2:00 271.6 F 29.3  
3:00 272.3 F 29.4  
4:00 271.4 F 29.0  
  
Min 271.4 F 28.9  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 220.7 F 0.0  
10:00 217.8 F 0.0  
15:00 206.8 F 0.0  
20:00 211.6 F 0.0  
25:00 227.5 F 0.0  
30:00 234.4 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

03/27/2023 03:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_