
Midmark M11 - v1.0.0
Total Cycles: 3570

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 182.8 F 2.7
4:00 197.7 F 6.7
6:00 225.8 F 11.3
8:00 227.1 F 5.8
10:00 235.8 F 10.8
12:00 235.5 F 7.3
14:00 245.3 F 13.1
16:00 244.5 F 12.4
18:00 256.7 F 18.8
20:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.6
2:00 272.5 F 29.0
3:00 271.8 F 28.3
4:00 272.7 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 221.1 F 0.0
10:00 234.4 F 0.0
15:00 228.8 F 0.0
20:00 225.3 F 0.0
25:00 221.4 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

07/24/2023 02:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____