
Midmark M11 - v1.0.5
Total Cycles: 2724

Name: RITTER2
Sterilizer ID: V2313228
08/02/2023 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.3 F 0.0
2:00 140.0 F 0.0
4:00 145.8 F 0.0
6:00 182.3 F 1.0
8:00 201.5 F 5.2
10:00 215.4 F 9.7
12:00 227.3 F 9.0
14:00 234.4 F 10.2
16:00 235.0 F 8.9
18:00 244.9 F 13.4
20:00 243.6 F 12.3
22:00 255.7 F 18.5
24:00 266.2 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.3
2:00 271.7 F 28.6
3:00 271.9 F 29.2
4:00 272.5 F 29.1

Min 271.2 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 206.1 F 0.0
10:00 216.5 F 0.0
15:00 217.6 F 0.0
20:00 214.4 F 0.0
25:00 217.4 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:50
CYCLE COMPLETE

08/02/2023 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____