

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2038

Name: RITTER 1  
Sterilizer ID: V2178198  
06/21/2022 07:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 98.4 F 0.0  
4:00 151.5 F 0.0  
6:00 172.8 F 0.2  
8:00 185.7 F 2.2  
10:00 207.4 F 4.7  
12:00 222.6 F 7.5  
14:00 232.4 F 10.9  
16:00 224.2 F 7.2  
18:00 234.4 F 9.1  
20:00 243.8 F 13.4  
22:00 238.7 F 10.0  
24:00 236.8 F 8.4  
26:00 247.5 F 13.7  
28:00 258.8 F 20.0  
30:00 269.1 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 271.5 F 28.6  
2:00 271.8 F 28.6  
3:00 272.4 F 28.8  
4:00 272.6 F 29.1

Min 271.1 F 28.6  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 221.3 F 0.0  
10:00 230.6 F 0.0  
15:00 210.1 F 0.0  
20:00 214.3 F 0.0  
25:00 217.5 F 0.0  
30:00 214.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:53  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:50  
CYCLE COMPLETE

06/21/2022 08:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02