
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000422
11:01 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:12
STERILIZING 3:15
136.0 C 224kPa 3:15
136.4 C 230kPa 3:45
136.4 C 233kPa 4:15
136.4 C 233kPa 4:45
136.3 C 233kPa 5:15
136.4 C 233kPa 5:45
136.4 C 233kPa 6:15
136.4 C 233kPa 6:45
136.4 C 233kPa 6:45
VENTING 6:45
AIR DRYING 7:39
CYCLE COMPLETE 7:46

Digital Signature #
5CE4AA15F6578DB2
