
Midmark M11 - v1.0.0
Total Cycles: 2739

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.3 F 0.0
2:00 83.4 F 0.0
4:00 88.0 F 0.0
6:00 116.7 F 0.0
8:00 150.2 F 0.0
10:00 179.0 F 1.7
12:00 195.1 F 4.8
14:00 214.7 F 8.2
16:00 226.0 F 12.0
18:00 224.7 F 6.3
20:00 235.1 F 10.9
22:00 233.2 F 7.4
24:00 244.2 F 13.1
26:00 242.1 F 11.4
28:00 254.1 F 17.5
30:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.9
2:00 272.3 F 29.1
3:00 271.9 F 28.5
4:00 272.7 F 29.2

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 204.0 F 0.0
10:00 194.9 F 0.0
15:00 199.2 F 0.0
20:00 197.5 F 0.0
25:00 205.6 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

01/05/2023 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____