
Midmark M11 - v1.0.5
Total Cycles: 1562

Name: RITTER2
Sterilizer ID: V2313228
10/04/2022 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 165.8 F 0.1
4:00 184.2 F 0.9
6:00 201.9 F 5.5
8:00 220.4 F 10.7
10:00 226.6 F 7.2
12:00 240.9 F 13.1
14:00 241.6 F 11.8
16:00 241.5 F 11.1
18:00 254.9 F 18.2
20:00 266.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.9 F 29.3
2:00 271.6 F 28.8
3:00 272.7 F 29.4
4:00 271.7 F 28.7

Min 271.6 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 214.5 F 0.0
10:00 236.2 F 0.0
15:00 231.9 F 0.0
20:00 220.7 F 0.0
25:00 213.8 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

10/04/2022 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____