
Midmark M11 - v1.0.0
Total Cycles: 3332

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2023 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 181.7 F 0.0
4:00 189.6 F 1.3
6:00 205.2 F 6.6
8:00 220.9 F 11.7
10:00 227.6 F 7.3
12:00 239.4 F 12.7
14:00 238.4 F 10.3
16:00 235.8 F 8.2
18:00 249.7 F 15.1
20:00 261.2 F 21.6
22:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 29.1
2:00 272.5 F 28.9
3:00 271.9 F 28.4
4:00 271.4 F 28.3

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 211.1 F 0.0
10:00 215.8 F 0.0
15:00 212.4 F 0.0
20:00 212.1 F 0.0
25:00 204.5 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:42
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

06/05/2023 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____