

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004185  
15:39 09/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 3:58  
136.0 C 220kPa 3:58  
136.4 C 225kPa 4:28  
136.4 C 224kPa 4:58  
136.2 C 222kPa 5:28  
136.4 C 225kPa 5:58  
136.2 C 222kPa 6:28  
136.2 C 224kPa 6:58  
136.3 C 224kPa 7:28  
136.3 C 224kPa 7:28  
VENTING 7:28  
AIR DRYING 8:23  
CYCLE COMPLETE 8:26

-----  
Digital Signature #  
2555899FB288D87A  
-----