

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001066  
13:18 30/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:37  
STERILIZING 3:39  
136.0 C 220kPa 3:39  
136.4 C 224kPa 4:09  
136.4 C 223kPa 4:39  
136.1 C 222kPa 5:09  
136.4 C 223kPa 5:39  
136.1 C 221kPa 6:09  
136.4 C 223kPa 6:39  
136.4 C 224kPa 7:09  
136.4 C 224kPa 7:09  
VENTING 7:09  
AIR DRYING 7:58  
CYCLE COMPLETE 8:00

-----  
Digital Signature #  
B55D85819BE3D825  
-----